



**Office of the Regional Executive Director**

**Regional Field Unit No. 10**

Antonio Luna St., Cagayan de Oro City

PABX (Telefax) No. (088-22) 725-725; (088) 856-2755

Email Address: [agri10@cagayandeoro.da.gov.ph](mailto:agri10@cagayandeoro.da.gov.ph)

**SUPPLEMENTAL/BID BULLETIN NO. 22**

May 3, 2013

This Supplemental/Bid Bulletin No. 22 is issued to modify items in the Bidding Documents for the Procurements of **1 UNIT MICROWAVE DIGESTER with ABC of Php 2,200,000.00**. This shall be an integral part of the Bidding Documents.

Please be informed of the following changes:

1. Minimum Specifications:

FROM : 1. Digital stainless steel cavity housing

TO : **1. Digital display**  
**2. Stainless steel cavity housing**

FROM : 2. Must have 10 places segmented rotor dual magnetron system and rotating diffuser equipped with safe view

TO : **3. With at least 10 places but not more than 16 places individually handled rotor for ease of handling of the analyst, dual magnetron system, equipped with homogenizer with safe view.**

FROM : 8. Power: 220 volts

TO : **8. Power: 220 volts; single-phased**

2. Delete minimum specifications #5. Sample motor must be able to operate with less 10 samples when required.

3. Opening of Bids

FROM : May 17, 2013; 10:00am

TO : **May 16, 2013; 9:00am**

**RTD CONSTANCIO C. MAGHANOY, Jr.**

Chair, Bids and Awards Committee

Received by the Bidder:

\_\_\_\_\_  
Signature Over Printed Name

Date: \_\_\_\_\_